


Fire Propagation of Products Test Equipment to BS 476 Part A1 2009



- ⇒ Electrical: 230 volts – Nominal 10 Amps
- ⇒ Ambient Temperature: Operating 10°C to 35°C
- ⇒ Dimensions: Apparatus: 400mm (W) x 600mm x (H) x 300mm (D)
- ⇒ Control module: 400mm (W) x 500mm x (H) x 400mm (D)
- ⇒ Gas Supplies: The gas supply specified in the Standard is Standard test gas G112, as specified in BS 4947.
- ⇒ Flow: Gas flow for burner adjustable from 0 to 5 nl/min
- ⇒ Pressure: 1kPa



E-41, Okhla Industrial Area, Phase-II
New Delhi-110020 (India)
Ph:- 91-11-40522000
Fax:- 91-11-40503150/51
www.dngtechno.com
E-mail:- enq@spectro.in
care@spectro.in