

Temperature Index Test Module



TECHNICAL SPECIFICATIONS:

- Electrical: 230vAC 50Hz / 115V 60Hz
- Operating Temperature: 10°C to 35°C
- Test Temperature: Ambient to 400°C
- Flow: Flow through column adjustable from 0 to 20 nl/min
- Gas Supplies: Oxygen and Nitrogen at 2.6 Bar (38 psi) minimum Minimum flow to Oxygen and Nitrogen pipe lines 25 nl/min (normally bottled supply with in-line regulator and filter). No other restrictions should be used as this will reduce the flow rates
- Propane at 0.1 – 0.5 Bar
- Oxygen Analyser: Range: 0 to 100% Oxygen
- Repeatability (typical): $\pm 0.1\%$ Oxygen
- Linearity (typical): $\pm 0.1\%$ Oxygen



E-41, Okhla Industrial Area, Phase-II
New Delhi-110020 (India)
Ph:- 91-11-40522000
Fax:- 91-11-40503150/51
www.dngtechno.com
E-mail:- enq@spectro.in
care@spectro.in